



Unimed S.A.  
Grand-Pré 10  
Case postale  
CH-1002 Lausanne  
Suisse - Switzerland

Téléphone +41 (0) 21 624 21 51  
Téléfax +41 (0) 21 624 53 32  
E-mail : info@unimed.ch  
Web : www.unimed.ch

**CLIENT**  
**City, Country**

**Control Plan**  
**EN 10204-3.1.B**

Customer	Article N° / Description :	Drawing N°:	Order N°:
	<b>9216621 Tube</b>	<b>9216621 / A</b>	<b>4333414</b>

Unimed	Reference:	Dimensions:	
	<b>200.010</b>	<b>0.80/0.60 x 1000 mm</b>	
	Order Reference N°:	Quantity:	Lot N°:
	<b>H 0506'456</b>	<b>22 pcs</b>	<b>0526</b>

Material	Description:							Tensile Strength [N/mm <sup>2</sup> ]:			
	<b>AISI 316L SHO 060103</b>							<b>1150</b>			
	Chemical Composition (%)	<b>C</b>	<b>Si</b>	<b>Mn</b>	<b>P</b>	<b>S</b>	<b>Cr</b>	<b>Mo</b>	<b>Ni</b>		
	<b>0.012</b>	<b>0.66</b>	<b>1.29</b>	<b>0.028</b>	<b>0.001</b>	<b>17.29</b>	<b>2.08</b>	<b>12.08</b>			

Control	Control Specifications :	Test results:
	<b>200010.PC</b>	<b>According to the enclosed control plan</b>

This document certifies that this product is in accordance with the specifications and drawings mentioned above.

Lausanne, April 27<sup>th</sup> 2006

UNIMED S.A.

**Claude Borgeaud**

C. Borgeaud  
Directeur AQ